

ed

Notice of Allowability

Application No.

10/686,910

Examiner

Leslie R. Deak

Applicant(s)

GINGGEN ET AL.

Art Unit

3761

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to amendment filed 15 March 2007.
2. ☒ The allowed claim(s) is/are 1,2 and 8-14.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

- | | |
|--|---|
| 1. <input type="checkbox"/> Notice of References Cited (PTO-892) | 5. <input type="checkbox"/> Notice of Informal Patent Application |
| 2. <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 6. <input type="checkbox"/> Interview Summary (PTO-413),
Paper No./Mail Date _____ |
| 3. <input type="checkbox"/> Information Disclosure Statements (PTO/SB/08),
Paper No./Mail Date _____ | 7. <input type="checkbox"/> Examiner's Amendment/Comment |
| 4. <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit
of Biological Material | 8. <input checked="" type="checkbox"/> Examiner's Statement of Reasons for Allowance |
| | 9. <input type="checkbox"/> Other _____ |

DETAILED ACTION

Allowable Subject Matter

1. Claims 1, 2, and 8-14 are allowed.
2. The following is an examiner's statement of reasons for allowance: The prior art fails to disclose or suggest the device and method claimed by applicant. In particular, the prior art fails to disclose an adjustable resistance valve for a cerebrospinal fluid shunt wherein the valve comprises a resistance system comprising a set of passages of identical diameter and varying length disposed in a helicoidal shape on a selection device, wherein the valve system is designed to select a single passage at a time, along with the other steps and limitations of the claims.

The best prior art of record, US 3,170,483 to Milroy and US 2005/0055009A1 to Rosenberg fail to disclose or suggest the device and method claimed by applicant. Rosenberg discloses an adjustable resistance valve for a cerebrospinal fluid shunt comprising a series of passages with varying diameter to provide varied resistance to flow. Milroy discloses a flow regulator valve with passages of varying length and cross-sectional area that provide varied resistance to flow, but the system is designed to select multiple passages at once and does not comprise a helicoidal shape. Accordingly, the claimed invention is patentable over the prior art.

US 6,095,491 to Kriesel discloses a cylindrical adjustable resistance valve with flow passages of identical diameter and identical length that comprise inserts with various flow resistance values. Kriesel fails to teach flow passages of varying length or a valve with a helicoidal shape as claimed by applicant.

Art Unit: 3761

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."


Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Leslie R. Deak whose telephone number is 571-272-4943. The examiner can normally be reached on M-F 7:30-5:00, every other Friday off.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Tanya Zalukaeva can be reached on 571-272-1115. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3761

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Leslie R. Deak
Patent Examiner
Art Unit 3761
26 April 2007

TATYANA ZALUKAEVA
SUPERVISORY PRIMARY EXAMINER

